

Member State  
P.R.China

OIML Certificate No.  
R60/2000-CN1-06.01

# OIML CERTIFICATE OF CONFORMITY

## Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine  
Address: N0.9 Madiandonglu, Haidian District, Beijing, 100088 China

## Applicant

Name: Zhonghang Electronic Measuring Instrument CO., LTD  
Address: P.O.Box2, hanzhong 723007, ShaanXi, China

## Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:

Address:

## Identification of the certified type

### Load Cell

Model designation	L 6 F
Accuracy class	C3
Maximum capacity ( $E_{max}$ )	50kg, 100kg, 150kg, 200kg, 250kg, 500kg, 750kg, 1000kg, 2000kg
Ratio of minimum LC verification interval $Y = E_{max}/V_{min}$	6000, 7500, 10000 or 15000
Minimum dead load ( $E_{min}$ )	0kg
Maximum number of load cell interval ( $n_{max}$ )	3000
Apportionment factor ( $P_{LC}$ )	0.7
Safe overload	150% $E_{max}$
Compensating Temperature range	-10 — +40

The numbers of the associated test reports and the included pages:

2005-WJ-0024 18 pages , 2004-WJ-0041 17 pages , 2004-WJ-0061 17pages

This certificate attests the conformity of the above-mentioned type(represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).

R: 60

Edition: 2000

For accuracy class (if applicable): C3

This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned ,as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval. The conformity was established by tests described in the associated test reports.

The issuing authority

Office of OIML Affairs

中国秘书处

March 28, 2006

The CIML member

王秦平

March 28, 2006

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.